

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property Organization
International Bureau



(43) International Publication Date
25 April 2002 (25.04.2002)

PCT

(10) International Publication Number
WO 02/32303 A2

(51) International Patent Classification⁷: **A61B 5/00**

[US/US]; 7197 Indian Valley Court, San Jose, CA 95139 (US).

(21) International Application Number: **PCT/US01/42570**

(74) Agents: **HAN, Johnney, U. et al.**; Morrison & Foerster, LLP, 755 Page Mill Road, Palo Alto CA 94304-1018 (US).

(22) International Filing Date: **9 October 2001 (09.10.2001)**

(25) Filing Language: **English**

(26) Publication Language: **English**

(30) Priority Data:
09/693,202 19 October 2000 (19.10.2000) **US**

(81) Designated States (*national*): AE, AG, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, BZ, CA, CH, CN, CO, CR, CU, CZ, DE, DK, DM, DZ, EC, EE, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MA, MD, MG, MK, MN, MW, MX, MZ, NO, NZ, PH, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, TZ, UA, UG, US, UZ, VN, YU, ZA, ZW.

(63) Related by continuation (CON) or continuation-in-part (CIP) to earlier application:
US 09/693,202 (CIP)
Filed on 19 October 2000 (19.10.2000)

(84) Designated States (*regional*): ARIPO patent (GH, GM, KE, LS, MW, MZ, SD, SL, SZ, TZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE, TR), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, ML, MR, NE, SN, TD, TG).

(71) Applicant (*for all designated States except US*): **MEDOPTIX, INC.** [US/US]; 10011 N. Foothill Blvd., Suite 107, Cupertino, CA 95014 (US).

Published:

— *without international search report and to be republished upon receipt of that report*

(72) Inventors; and

(75) Inventors/Applicants (*for US only*): **BERMAN, Herbert, L.** [US/US]; 12680 Viscaino Road, Los Altos Hills, CA 94022 (US). **ROE, Jeffrey, N.** [US/US]; 3212 Veracruz Drive, San Ramon, CA 94583 (US). **BLAIR, Robert, N.**

For two-letter codes and other abbreviations, refer to the "Guidance Notes on Codes and Abbreviations" appearing at the beginning of each regular issue of the PCT Gazette.

WO 02/32303 A2

(54) Title: **GLUCOSE MEASUREMENT UTILIZING NON-INVASIVE ASSESSMENT METHODS**

(57) Abstract: This involves non-invasive glucose measurement processes for determining blood glucose level in the human body. After achieving a static level of glucose at a surface of the skin over some period of time, the glucose may then be measured by a variety of different processes. A sample of the glucose may also first be extracted from the skin and this sample may then be measured. Clearly, these processes are especially suitable for monitoring glucose levels in the human body, and is especially beneficial to users having diabetes mellitus. These procedures may be used for other analyte materials that are found in appropriate regions of the outerskin.

GLUCOSE MEASUREMENT UTILIZING NON-INVASIVE ASSESSMENT METHODS

Field of the Invention

5 This invention involves non-invasive glucose measurement and a process for determining blood glucose levels in the human body. Preferably, the process is used on a fingertip or other part of the body, typically a skin surface of the body.

Background of the Invention

10 The American Diabetes Association reports that nearly 6% of the population in the United States, a group of 16 million people, has diabetes. The Association further reports that diabetes is the seventh leading cause of death in the United States, contributing to nearly 200,000 deaths per year. Diabetes is a chronic disease having no cure. The complications of the disease include blindness, kidney disease, nerve disease, and heart disease, perhaps with stroke. Diabetes is said to be the leading cause of new cases of
15 blindness in individuals in the range of ages between 20 and 74; from 12,000-24,000 people per year lose their sight because of diabetes. Diabetes is the leading cause of end-stage renal disease, accounting for nearly 40% of new cases. Nearly 60-70% of people with diabetes have mild to severe forms of diabetic nerve damage which, in severe forms, can lead to lower limb amputations. People with diabetes are 2-4 times more likely to have
20 heart disease and to suffer strokes.

Diabetes is a disease in which the body does not produce or properly use insulin, a hormone needed to convert sugar, starches, and the like into energy. Although the cause of diabetes is not completely understood, genetics, environmental factors, and viral causes have been partially identified.

25 There are two major types of diabetes: Type I and Type II. Type I diabetes (formerly known as juvenile diabetes) is an autoimmune disease in which the body does not produce any insulin and most often occurs in young adults and children. People with Type I diabetes must take daily insulin injections to stay alive.

Type II diabetes is a metabolic disorder resulting from the body's inability to make
30 enough, or properly to use, insulin. Type II diabetes accounts for 90-95% of diabetes. In the United States, Type II diabetes is nearing epidemic proportions, principally due to an

increased number of older Americans and a greater prevalence of obesity and a sedentary lifestyle.

Insulin, in simple terms, is the hormone that unlocks the cells of the body, allowing glucose to enter those cells and feed them. Since, in diabetics, glucose cannot enter the
5 cells, the glucose builds up in the blood and the body's cells literally starve to death.

Diabetics having Type I diabetes typically are required to self-administer insulin using, e.g., a syringe or a pin with needle and cartridge. Continuous subcutaneous insulin infusion via implanted pumps is also available. Insulin itself is typically obtained from
10 pork pancreas or is made chemically identical to human insulin by recombinant DNA technology or by chemical modification of pork insulin. Although there are a variety of different insulins for rapid-, short-, intermediate-, and long-acting forms that may be used variously, separately or mixed in the same syringe, use of insulin for treatment of diabetes is not to be ignored.

It is highly recommended by the medical profession that insulin-using patients
15 practice self-monitoring of blood glucose (SMBG). Based upon the level of glucose in the blood, individuals may make insulin dosage adjustments before injection. Adjustments are necessary since blood glucose levels vary day to day for a variety of reasons, e.g., exercise, stress, rates of food absorption, types of food, hormonal changes (pregnancy, puberty, etc.) and the like. Despite the importance of SMBG, several studies have found that the
20 proportion of individuals who self-monitor at least once a day significantly declines with age. This decrease is likely due simply to the fact that the typical, most widely used, method of SMBG involves obtaining blood from a finger stick. Many patients consider obtaining blood to be significantly more painful than the self-administration of insulin.

There is a desire for a less invasive method of glucose measurement. Methods exist
25 or are being developed for a minimally invasive glucose monitoring, which use body fluids other than blood (e.g., sweat or saliva), subcutaneous tissue, or blood measured less invasively. Sweat and saliva are relatively easy to obtain, but their glucose concentration appears to lag in time significantly behind that of blood glucose. Measures to increase sweating have been developed and seem to increase the timeliness of the sweat glucose
30 measurement, however.

Subcutaneous glucose measurements seem to lag only a few minutes behind directly measured blood glucose and may actually be a better measurement of the critical

values of glucose concentrations in the brain, muscle, and in other tissue. Glucose may be measured by non-invasive or minimally-invasive techniques, such as those making the skin or mucous membranes permeable to glucose or those placing a reporter molecule in the subcutaneous tissue. Needle-type sensors have been improved in accuracy, size, and stability and may be placed in the subcutaneous tissue or peripheral veins to monitor blood glucose with small instruments. See, "*An Overview of Minimally Invasive Technologies*", Clin. Chem. 1992 Sep.; 38(9):1596-1600.

Truly simple, non-invasive methods of measuring glucose are not commercially available.

SUMMARY OF THE INVENTION

This invention involves non-invasive glucose measurement and a process for determining blood glucose levels in the human body upon achieving a static level of glucose at a skin surface over a period of time.

Processes which are able to assess glucose concentrations predictably from a skin surface may include a step of extracting a sample from the skin and then measuring that sample from the skin. Such sample extraction processes may include suction blister extraction, wick extraction, microdialysis extraction, iontophoretic extraction, sonophoretic extraction, and chemically enhanced extraction. Aside from the extraction processes, non-invasive measurement processes may include electrochemical sensors (e.g., glucose electrodes), optochemical sensors (e.g., colorimetric strips), near-infrared spectroscopy (NIR), mid-infrared spectroscopy (MIR), infrared spectroscopy (IR), Raman spectroscopy, photoacoustic spectroscopy, measurement of refractive index or scatter changes, fluorescent spectroscopy, and polarization spectroscopy.

The processes for extraction and measurement are illustrative and are not meant to be an exclusive list.

BRIEF DESCRIPTION OF THE DRAWINGS

Figure 1 shows a graph representing a glucose level rising sharply and reaching a static value.

Figure 2 shows a graph correlating glucose levels measured using a specific variation of the device with glucose levels in the blood determined using a commercial device.

Figure 3 shows a pair of glucose IR curves (taken before and after eating) for an individual having diabetes made using the inventive glucose measuring device.

Figure 4 shows a graph comparing glucose levels in a non-diabetic individual (taken before and after eating) made using the inventive glucose measuring device and direct blood measurement. This graph shows that the inventive procedure tracks blood glucose levels with minimum time lag.

DESCRIPTION OF THE INVENTION

In non-invasively measuring analyte levels using mid-infrared ("MIR"), particularly using a device and methods as described in U.S. Patent Application Serial No. 09/547,433, filed April 12, 2000, entitled "INFRARED ATR GLUCOSE MEASUREMENT SYSTEM (II)", which is herein incorporated by reference in its entirety, penetration of skin by IR ranged in only a few micrometers. Thus, because of this small penetration depth, the described device may have been measuring glucose from the mixture of oils and sweat that is pumped to the skin surface. Therefore, multiple alternative processes may be utilized in non-invasively measuring analytes, particularly glucose, from a skin surface.

The device in co-pending Application Serial No. 09/547,433 described the use of IR attenuated total reflectance ("ATR") spectroscopy to detect and ultimately to determine the level of a selected analyte, preferably blood glucose, in the human body. Preferably, the inventive device used an ATR procedure in which the size and configuration of the crystal permits a number of internal reflections before the beam is allowed to exit the crystal with its measured information.

We have found that the mid-IR spectrum does not penetrate into the skin to an appreciable level. Specifically, the skin is made up of a number of layers: the outermost -- the *stratum corneum* -- is a layer substantially free of interference from cholesterol, water, gamma globulin, albumin, and blood. It is a shallow outer region covering the *stratum granulosum*, the *stratum spinosum*, and the basal layer. The area between the basal layer to the outside is not vascularized. It is unlikely that any layer other than the *stratum corneum* is traversed by the mid-IR light involved in this inventive device. Although we do not wish to be bound by theory, it is likely that the eccrine or sweat glands transport the glucose to the outer skin layers for measurement and analysis by our inventions.

We have surprisingly found that a glucose measuring device made according to the invention in Application Serial No. 09/547,433 is quite effective on the human skin of the hands and fingers. We have found that the glucose concentration as measured by the inventive devices correlates very closely with the glucose concentration determined by a direct determination from a blood sample. This is surprising in that the IR beam likely passes into the skin, i.e., the *stratum corneum*, for only a few microns. It is unlikely in a fingertip that any blood is crossed by that light path. As discussed above, the *stratum corneum* is the outer layer of skin and is substantially unvascularized. The *stratum corneum* is the final outer product of epidermal differentiation or keratinization. It is made up of a number of closely packed layers of flattened polyhedral corneocytes (also known as squames). These cells overlap and interlock with neighboring cells by ridges and grooves. In the thin skin of the human body, this layer may be only a few cells deep, but in thicker skin, such as may be found on the toes and feet, it may be more than 50 cells deep. The plasma membrane of the corneocyte appears thickened compared with that of keratinocytes in the lower layers of the skin, but this apparent deposition of a dense marginal band formed by stabilization of a soluble precursor, involucrin, just below the *stratum corneum*.

It is sometimes necessary to clean the skin exterior prior before sampling to remove extraneous glucose from the skin surface. At least when using IR spectra to measure glucose, it is important to select cleaning materials having IR spectra that do not interfere with the IR spectra of glucose. We consider a kit of the following to be suitable for preparation of the sample skin for the testing. The components are: a.) a glucose solvent, e.g., water or other highly polar solvent; b.) a solvent for removing the water, e.g., isopropanol, and c.) a skin softener or pliability enhancer not having significant IR peaks in the noted IR regions, e.g., mineral oils such as those sold as "Nujol". Preferably the b.) and c.) components are admixed, although they need not be. Certain mixtures of the first two components may be acceptable, but only if the sampling situation is such that the solvents evaporate without IR spectrographically significant residue. We have also found that soap and its residue are sometimes a problem. Consequently, addition of a weak acid again not having significant IR peaks in the noted IR regions, to the a.) component, i.e., the solvent for removing glucose, is desirable. The preferred weak acid is boric acid. The inventive kit preferably is made up of sealed packets of the components, most preferably each packet containing an absorbent pad.

Method of Use

As noted in Application Serial No. 09/547,433, for IR measurement, it is desirable both to clean the plate before use and to clean the exterior surface of the skin to be
5 sampled. Again, we have found that the exterior skin is highly loaded with glucose that is easily removed preferably by using the skin preparation kit, or, less preferably, by washing the skin. Reproducible and accurate glucose measurements may then be had in a period as short as one to ten minutes, generally less than five minutes, after cleaning the area of the skin to be measured.

10 We also note that, depending upon the design of a specific variation of a device made according to the invention, periodic at least an initial calibration of the device, using typical blood sample glucose determinations, may be necessary or desirable.

Alternate Methods of Use

15 We have also observed the following phenomena. In attempting to measure glucose using the mid IR apparatus described above, we noted that by using a solvent on the surface which dissolved glucose, our procedures showed that the glucose level at the skin was substantially lowered, if not eliminated. However, as is shown in the examples, shortly thereafter, the glucose level began to rise sharply and consequently would reach static value
20 which was correlateable in a consistent fashion to a glucose level found in the blood.

As shown in Figure 1, the curve representing the glucose level rises sharply and eventually plateaus over a time period, T. This glucose level may be measured by any of the devices and processes as discussed in Application Serial No. 09/547,433 or herein at predetermined time periods, ΔT , until the glucose level reaches a substantially static value,
25 thus representing the glucose level in the blood. As we have noted above, the outer layer of the skin is not vascularized and the physiological source of the glucose transport to the skin surface is not all together clear. Nevertheless, it is easily assessable and quite repeatable. Our method of using mid IR to measure this glucose level is believed, on basic principals, only to penetrate the skin at best a few microns. Other procedures which are non-invasive
30 and which are able to assess glucose concentration upon this achievement of glucose stasis level are similarly and predictably able to assess the glucose level in the human body.

Such alternative procedures may provide for extracting a sample from the skin and then measuring the analyte, or glucose, level in the sample. Potential extraction methods may include blister suction and wick extraction, microdialysis extraction, iontophoretic extraction, iontophoretic extraction, and chemically enhanced extraction.

5 Sample measurement methods may include electrochemical sensors (e.g., glucose electrodes), optochemical sensors (e.g., colorimetric strips), near-infrared spectroscopy (NIR), infrared spectroscopy (IR), Raman spectroscopy, photoacoustic spectroscopy, measurement of refractive index or scatter changes, fluorescent spectroscopy, and polarization spectroscopy. These methods of sample extraction and measurement are
10 provided

Glucose Extraction

Blister suction and wick extraction are some of the most common methods for sampling subcutaneous interstitial tissue fluid, although blister extraction is less invasive
15 than the wick extraction technique. Microdialysis extraction involves calculating the concentrations of compounds, including skin glucose concentrations, which are in the extracellular water space. Microdialysis has been applied to peripheral tissue types, e.g., skin, muscle, adipose, eye, lung, liver, and blood as well as having microdialysis probes implanted subcutaneously and perfused by a portable microinfusion pump. Finally,
20 iontophoretic extraction involves noninvasive glucose measurement from subcutaneous tissue.

Glucose Measurement

25 Electrochemical Sensors

Electrochemical sensors (e.g., glucose electrodes) utilize electrical signal as a direct consequence of some (chemical) process occurring at a transducer/analyte interface. Some implantable glucose sensors may include electrocatalytic sensors, which are based on direct electro-oxidation of glucose on noble metal electrodes, and biosensors, which combine
30 glucose-specific enzymes with electrochemical electrodes.

Such sensors may be fabricated by combining biologically active components (e.g., enzymes, antibodies, cells, tissues or microorganisms) with some physical transducer.

Biosensors may be direct enzyme biosensors or affinity sensors based on enzyme labeled immunoassays. Enzymes may be used as a molecular recognition element in glucose sensing while immunoassays may provide the ability to sense extremely low amounts of an analyte. Electrochemical sensors may also include piezoelectric, thermoelectric, and acoustic sensors used for glucose measurement by utilizing an enzyme-catalysed reaction to create a measurable change in a physical parameter detectable by a transducer.

Optochemical Sensors

Optochemical sensors (e.g., colorimetric strips) are based on changes in some optical parameter due to enzyme reactions or antibody-antigen bonding at a transducer interface. Such sensors may include enzyme optrodes and optical immunosensors and may also include different monitoring processes such as densitometric, refractometric or colorimetric devices.

Electrochemical biosensors may be constructed on the amperometric principle which is based on the oxidation or reduction of electrochemically active substances. Such sensor may also be constructed to measure the changes in local pH due to the gluconic acid produced at a potentiometric sensor, usually a coated wire pH-selective electrode or an ion selective field effect transistor (ISFET). Also, electrical resistance changes during the overall process may be used as a basis for conductometric biosensors.

Moreover, potentiometric glucose sensors (e.g., coated wire sensors) may potentially be utilized for implantable use. Coated wire sensors are general easy to fabricate and are suitable for miniaturization to diameters of 50-200 μm . They may also be used in combination with a standard cardiographic (EKG) reference electrode.

Near-Infrared Spectroscopy (NIR)

The NIR region of the spectrum extends from 700 to 2500 nm ($14,000\text{--}4000\text{ cm}^{-1}$). In this region, absorption bands are due to overtone vibrations of anharmonic fundamental absorption bands to combinations of fundamental absorption bands primarily associated with C-H, O-H, and N-H stretching vibrations. For overtone vibrations, only the first, second, and third overtones are usually seen with the magnitude of the absorption peak diminishing substantially with overtone order. The NIR region may be attractive for quantitative spectroscopy since NIR instrumentation is readily available.

In measuring aqueous glucose, the NIR region which lies between 2.0 and 2.5 μm may be utilized. This region contains a relative minimum in the water absorption spectrum and has readily identifiable glucose peak information. However, NIR spectra may generally be sensitive to a host of factors including temperature, pH, and scattering.

5

Raman Spectroscopy

Raman spectra are typically observed when incident light at a frequency $\nu_0 = c/\lambda_0$ is inelastically scattered at frequencies $\nu_0 \pm \nu_i$. The loss (Stokes shift) or gain (anti-Stokes shift) of photon energy, and hence frequency, is due to transitions of the rotational and vibrational energy states within the scattering molecule. Since the Raman spectrum is independent of excitation frequency, an excitation frequency may be chosen which is appropriate for a particular sample. However, a drawback may be that scatter and reabsorption in biological tissues may make detection of Raman shifts due to physiological concentrations difficult.

10
15

Photoacoustic Laser Spectroscopy

Photoacoustic laser spectroscopy has been utilized for measuring glucose concentrations of human whole blood samples using pulsed laser photoacoustic spectroscopy. Such a process may use, e.g., a CO_2 laser operating with μJ pulse energy, to measure tiny changes of the absorption coefficient of the sample caused by the variations of blood glucose concentrations.

20

Refractive Index or Scatter Changes

Measurement of refractive index or scatter changes may be feasible to measure blood glucose by measuring the scattering coefficient of human skin, e.g., by using optical sensors attached to the skin. Such techniques may be based on the fact that the refractive index of sugar solution changes with the concentration of sugar.

25

Fluorescent Spectroscopy

There may be two categories for fluorescent spectroscopy: glucose-oxidase based sensors and affinity-binding sensors. Sensors in the first category may use the electroenzymatic oxidation of glucose by glucose-oxidase (GOX) in order to generate an

30

optically detectable glucose-dependent signal. The oxidation of glucose and oxygen forms gluconolactone and hydrogen peroxide.

Several methods for optically detecting the products of this reaction, and hence the concentration of glucose driving the reaction, may be utilized. Since oxygen is consumed in this reaction at a rate dependent on the local concentration of glucose, a fluorophore which is sensitive to local oxygen concentration can also be used to quantify glucose concentration.

A method GOX based fluorescent sensor involves the redox mediator tetrathiafulvalene (TTF) whose oxidized form TTF^+ reacts with the reduced form of GOX to reversibly form TTF^0 . Since TTF^+ is absorbed in the 540-580 nm range, a method for quantifying the presence of TTF^+ (and hence glucose driving the production of reduced GOX) is available.

Another method involves the hydrogen peroxide (H_2O_2) generated from the GOX reaction with glucose reacting with bis(2,4,6-trichlorophenyl) oxalate (TCPO) to form a peroxyoxylate. Here, the peroxyoxylate formed transfers chemiluminescent energy to an accepting fluorophore which in turn emits photons at a characteristic wavelength. The emission by the fluorophore is proportional to the glucose concentration and may be detected optically.

Polarization Spectroscopy

Polarimetric quantification of glucose may be based on the principle of optical rotary dispersion (ORD) where a chiral molecule in an aqueous solution rotates the plane of linearly polarized light passing through the solution. This rotation is due to a difference in the indices of refraction n_L and n_R for left- and right-circularly polarized light passing through a solution containing the molecule. Because the molecule has a chirality (or "handedness"), the angle of rotation depends linearly on the concentration of the chiral species, the path length through the sample, and a constant for the molecule called the specific rotation. Glucose in the body is dextrorotatory, i.e., rotates light in a right-handed direction, and has a specific rotation of $+52.6^\circ \text{ dm}^{-1} (\text{g/L})^{-1}$.

EXAMPLES OF MID-IR USE

Example 1

Using a commercially available IR spectrometer (Nicolet 510) having a ZnSe
5 crystal ATR plate (55mm long, 10mm wide, and 4mm thick) we tested the inventive
procedure. We calibrated the output of the spectrometer by comparing the IR signal to the
values actually measured using one of the inventor's blood samples. The inventor used a
blood stick known as "Whisper Soft" by Amira Medical Co. and "Glucometer Elite" blood
glucose test strips sold by Bayer Corp. of Elkhart, Ind. On each of the various test days, the
10 inventor took several test sticks and measured the glucose value of the resulting blood; the
IR test was made at the same approximate time.

As shown in the calibration curve of Figure 2, the data are quite consistent. So,
where the blood glucose concentration "B" is in (mg/dl) and "S" is the difference between
the absorbance at the referencing region and the measuring region as measured by the
15 spectrometer:

$$B=[(1950) \bullet S]-(17).$$

Example 2

20 In accordance with a clinical protocol, a diabetic was then tested. Curve 1 in Figure
3 shows the IR absorbance spectrum of the test subject's finger before eating (and after
fasting overnight) and curve 2 shows IR absorbance spectrum of the same individual after
having eaten. Incidentally, insulin was administered shortly after the measurement of curve
2.

25 In any event, the significant difference in the two peak heights at the 9.75
micrometer wavelength and the equality of the two IR absorbance values at the 8.50
micrometer value shows the effectiveness of the procedure in measuring glucose level.

Example 3

30 That the inventive glucose monitoring device non-invasively determines blood
glucose level and quickly follows changes in that blood glucose level is shown in Figure 4.
Using both the inventive procedure and a commercial glucose device, one of the inventors

followed his glucose level for a single day. The blood sticks are considered to be accurate within 15% of the actual reading.

5 The results are shown in Figure 4. Of particular interest is the measurement just before 4:40pm wherein the two values are essentially the same. A high sugar candy bar was eaten at about 4:45pm and measurements of glucose level were taken using the inventive procedure at about 5:03, 5:18, 5:35 and 5:50. A blood sample was taken at 5:35 and reflected almost the same value as that measured using the inventive procedure. Consequently, the procedure tracks that measured by the blood very quickly.

10 This invention has been described and specific examples of the invention have been portrayed. The use of those specifics is not intended to limit the invention in any way. Additionally, to the extent there are variations of the invention with are within the spirit of the disclosure and yet are equivalent to the inventions found in the claims, it is our intent that this patent will cover those variations as well.

WE CLAIM AS OUR INVENTION:

1. A process for determining the glucose level of a human body, comprising the steps of:

- 5 a.) achieving a static level of glucose at the skin surface as measured over a period of time, and
 b.) measuring said skin surface glucose level.

2. The process of claim 1, wherein the step of measuring further comprises:

- 10 a.) extracting a sample from the skin surface. and
 b.) measuring the glucose level from the sample.

3. The process of claim 2, wherein the step of extracting is selected from the group consisting of blister suction, wick extraction, microdialysis extraction, iontophoretic
15 extraction, sonophoretic extraction, and chemically enhanced extraction.

4. The process of claim 1, wherein the step of measuring comprises electrochemical sensors.

- 20 5. The process of claim 4, wherein the electrochemical sensors comprise glucose electrodes.

6. The process of claim 1, wherein the step of measuring comprises optochemical sensors.

- 25 7. The process of claim 6, wherein the optochemical sensors comprise colorimetric strips.

8. The process of claim 1, wherein the step of measuring comprises near-infrared
30 spectroscopy.

9. The process of claim 1, wherein the step of measuring comprises Raman spectroscopy.

5 10. The process of claim 1, wherein the step of measuring comprises photoacoustic spectroscopy.

11. The process of claim 1, wherein the step of measuring comprises measuring a refractive index or scatter changes of the sample.

10 12. The process of claim 1, wherein the step of measuring comprises fluorescent spectroscopy.

13. The process of claim 1, wherein the step of measuring comprises polarization spectroscopy.
15




FIG. 1




FIG. 2




FIG. 3




FIG. 4